

QA Dental Engineers
Suction Maintenance and Service schedule

Details	
Practice Name	Dr A Filing
Address	Your Practice Lincoln Lincoln, LN1 1AB
Phone Number	
Compressor:-	
Model	Cattani
Serial No.	
Date of Manufacture	
Location	Front Surgery

Component Check Year 1	Visual	Physical	Comments (Action taken or Required)
Intake filter/s	✓	✓	Satisfactory
Check tubing's	✓	✓	Satisfactory
Check cables	✓	✓	Satisfactory
Check for leaks	✓	✓	None found
Check separator orifice plate	✓	✓	Replaced *
Check condition amalgam canister & O-ring	✓	✓	Replaced *
Non Return Valve	✓	✓	Satisfactory
Water draw off valve	✓	✓	Satisfactory
Check general operation	✓	✓	Satisfactory
Operating noise level (unusual or excessive)	✓	✓	Satisfactory
Seperator start capacitor	✓	✓	Replaced *
Waste outlet valve	✓	✓	Replaced *

	Sterilising Cycle		Units
	Run on time	30	Sec
	Flow rate at tip	20,000	l/min
	Static Pressure	150	mbar

Comment/Recommendations		
* Parts in service kit (fitted to pump)	Date	24/07/09
	Engineer (in Block)	BRIAN BACON
	Signature	
	Practice (in Block)	
	Signature	