

QA Dental Engineers Autoclave Service Record

Details			
Practice Name	Practice Name		
Address	Address 1		
	Address 2		
	Postcode		
Phone Number			
Autoclave:-			
Manufacturer	Eschmann		
Model	SES 2000		
Serial No.	SCB-9G-2304		
Date of Manufacture	1999		
Water Quality & No. of Cycles	10538/2PPM		
Service Level	SL2 1/2		
Component Check	Visual	Physical	Comments (action taken or required)
Door Hinges & Brackets	✓	✓	Checked
Door Handle	✓	✓	Checked
Door Sliding Gear	n	a	
Door Lock	✓	✓	Checked
Door Seal	✓	✓	New fitted
Front Cover	✓	✓	Cleaned
Tray Holder	✓	✓	Cleaned
Chamber Face	✓	✓	Cleaned
Chamber Interior	✓	✓	Cleaned
Heating Element (inc. washer)	✓	✓	Satisfactory
Water Sensor (inc. washer)	✓	✓	New washers fitted
Reservoir	✓	✓	Cleaned
Water Tensioner	✓	✓	Cleaned
Pipe work	✓	✓	Checked
Filters	✓	✓	Semi blocked
Solenoid Valve(s)	✓	✓	Stripped and inspected
Joints & Seals	✓	✓	Satisfactory
Electrical Connections	✓		Satisfactory
Door Interlocks - Electrical	✓	✓	Satisfactory
Door Interlocks - Mechanical	✓	✓	Satisfactory
Door Interlocks - Pressure	✓	✓	Satisfactory
Air Vent Valve	✓	✓	Working
Pressure Relief Valve	✓	✓	Tested during PSI
Calibration	✓	✓	Carried out - satisfactory
Schematic	Sterilising Cycle		Units
	Fill Time	1:19	min
	Heat Time	3:20	min
	Sterilising Time	3:15	min.
	Condensing Time	1:23	min.
	Overall Time	9:17	min.
	Sterilising Temp (min.)	135.6	deg C
	Sterilising Temp (max.)	135.8	deg C
	Sterilising Pressures (max/min)	2.2/2.15	Bar
Comment/Recommendations	Date	DATE	
	Engineer (in Block)	ENGINEER NAME	
	Signature		
	Practice (in Block)		
	Signature		